



## PureLife ESSENCE Latex Powder-Free Examination Glove

### Product Specification Sheet

Reorder #	Size : XS S M L XL	
Product Part #	<b>TBA</b>	
510K # / MDL #	K991044 / D026930	
Design and Feature	Ambidextrous, non-textured surface (ES) and beaded cuff	
Type	Powder free and non-sterile examination gloves	
Material	Natural rubber latex	
Color	Off-white to slight yellow	
Surface Treatment	Chlorination and polymer rinsed (2G-CP)	
Protein Content	This latex glove contains 50 micrograms or less of total water extractable protein per gram	
Powder Free Residue (mg/glove)	Maximum2	
Dimension - Palm Width (mm)	Size : XS S M L XL	75 ± 10 80 ± 10 95 ± 10 110 ± 10 115 ± 10
Dimension - Length (mm)	Size : XS S M L XL	Min. 230 Min. 230 Min. 230 Min. 230 Min. 230
Single-wall Thickness (mm) *all sizes	Finger Palm Cuff	Min. 0.13 Min. 0.12 Min. 0.10
Physical Properties (Before Aging)	Tensile Strength (MPa) Elongation (%) Modulus 500% (MPa)	Min. 18 Min. 650 Max. 5.5
Physical Properties (After Aging)	Tensile Strength (MPa) Elongation (%) Modulus 500% (MPa)	Min. 14 Min. 500 Not applicable
Packing Mode	100 gloves (by weight) per dispenser and 20 dispensers per carton. *Tolerance: ±2 gloves per dispenser	
Glove Marking	No marking	
Lot # identification on Finished Goods	YMMPPPPSS Y = Year of packing (eg. 9=year 2009, 0=year 2010, etc) MM = Month of packing (eg. 01=Jan, 02=Feb, etc) PPPP = The last 4 digit of WRP's packing work order # SS = glove size (i.e. 00=XS, 01=S, 02=M, 03=L and 04=XL)	
Pre-shipment Inspection	Dimension Physical Properties 1000ml Water Leak Protein Content Powder Free Residue	S-2, AQL 4.0 S-2, AQL 4.0 G-I, AQL 1.5 N=3 N=5
Product Conformance	ASTM D3578	