

PureLife SERENITY Latex Lightly Powdered Examination Glove

Product Specification Sheet

1. Product : Powdered Examination Gloves
2. Material : Natural Rubber Latex
3. Type : Non Sterile
4. Glove Design & Feature : Ambidextrous, Microtextured or Smooth Surface
5. Cuff : Beaded
6. Usage : For Single Use Only
7. Colour : Beige
8. Packaging : 100 ± 2 pcs/inner, 10 or 20 inners/carton
9. Dimensions (Available in 5 sizes) :

Size	Palm Width (mm)	SGM Length (mm)	ASTM D3578-05 Length (mm)
XS	70 – 79	Min. 240	Min. 220
S	80 – 89	Min. 240	Min. 220
M	90 – 99	Min. 240	Min. 230
L	100 – 109	Min. 240	Min. 230
XL	110 – 119	Min. 240	-
10. Thickness :

Location of Thickness Measurement	Single Wall (mm)
Finger	Min. 0.15
Palm	Min. 0.12
Cuff	Min. 0.08
11. Physical Properties :

	Before Aging	After Aging
Tensile Strength (MPa)	Min. 18	Min. 14
Elongation (%)	Min. 650	Min. 500
Modulus (500%)	Max. 5.5 MPa	-
Force At Break (N)	Min. 9	Min. 6
12. Protein Content : Max. 200 µg/dm²
13. Powder Level : Max. 10 mg/ dm²
14. Compliance Standards : ASTM D3578-05, EN 455 Part 1,2 & 3:2000
15. Quality Assurance : ISO 9001 : 2000 certified by SGS (UK) and SIRIM (Malaysia)
ISO 13485 : 2003 certified by SGS (UK)
CMDCAS ISO 13485 certification
Directive 93/42/EEC Annex V (sterility aspects only) – for powdered latex examination gloves